

## Specification

Name	CPAP		AutoCPAP		BiPAP	
Model	CW-CP1	CW-CP2	CW-AC8	CW-AC9	CW-BP8	CW-BP9
Display	3.5" TFT LCD	128*64 LCD	3.5" TFT LCD	128*64 LCD	3.5" TFT LCD	128*64 LCD
<b>Working Mode</b>						
CPAP	•	•	•	•	•	•
CPAP+EPR	•	•	•	•	•	•
Auto CPAP			•	•		
Titration			•	•		
S					•	
T					•	
S/T					•	•
<b>Features</b>						
Pressure (in 0.5cmH <sub>2</sub> O increment)	4-20cmH <sub>2</sub> O	4-20cmH <sub>2</sub> O	4-20cmH <sub>2</sub> O	4-20cmH <sub>2</sub> O	4-20cmH <sub>2</sub> O	4-20cmH <sub>2</sub> O
IPAP (in 0.5cmH <sub>2</sub> O increment)					4-20cmH <sub>2</sub> O	4-20cmH <sub>2</sub> O
EPAP (in 0.5cmH <sub>2</sub> O increment)					4-20cmH <sub>2</sub> O	4-20cmH <sub>2</sub> O
EPR (1-3 level)	2-4cmH <sub>2</sub> O	2-4cmH <sub>2</sub> O	2-4cmH <sub>2</sub> O	2-4cmH <sub>2</sub> O	2-4cmH <sub>2</sub> O	2-4cmH <sub>2</sub> O
Ramp Duration	0-60mins	0-60mins	0-60mins	0-60mins	0-60mins	0-60mins
Inspiration Sensitivity (1~6 level)					•	•
Expiration Sensitivity (1~6 level)					•	•
Respiratory Rate (4~40BPM)					•	•
Inspiration Time (0.5~3.0s)					•	•
Automatic ON/OFF	•	•	•	•	•	•
Automatic Leakage Compensation	•	•	•	•	•	•
Automatic Altitude Compensation	•	•	•	•	•	•
<b>Data Download</b>						
Data in Total (Artificial Intelligence, Leakage, Data Analysis, Adaptability, Usage Time)			•	•	•	•
Detail Data (AHI, Event, Pressure and Respiratory Airflow)	•	•	•	•	•	•
USB Download to PC	•	•	•	•	•	•
<b>Humidifier</b>						
Ajustable Constant Temperature Humidification System	•	•	•	•	•	•
Replaceable Water Chamber	•	•	•	•	•	•
<b>Other Parameters</b>						
Input Volt. & Freq. (110-240Vac, 50/60Hz)	•	•	•	•	•	•
Working dB (Sound Pressure Level)	<30dBA (working at 10cmH <sub>2</sub> O)					
Net Weight & Size	1.8kg, 255*170*112mm (Host+Humidifier)					

