


## Quick Guide

### 1. Power on




### 2. Add sample(10ul blood) into the hemolytic(50ul) and mix them to standby



### 3. Swipe card



### 4. After reading the card, press "Test" button and the screen will show as below



### 5. Insert new cartridge



### 6. Press start, drop 3ul sample to the right position of cartridge and push it back into the instrument



### 7. Drop the eluent B(35ul) to the left position



### 8. Test result display on the screen after 200 seconds

